

University of Groningen

## The influence of the sample matrix on LC-MS/MS method development and analytical performance

Koster, Remco Arjan

**IMPORTANT NOTE:** You are advised to consult the publisher's version (publisher's PDF) if you wish to cite from it. Please check the document version below.

*Document Version*

Publisher's PDF, also known as Version of record

*Publication date:*

2015

[Link to publication in University of Groningen/UMCG research database](#)

*Citation for published version (APA):*

Koster, R. A. (2015). *The influence of the sample matrix on LC-MS/MS method development and analytical performance*. [Thesis fully internal (DIV), University of Groningen]. University of Groningen.

### Copyright

Other than for strictly personal use, it is not permitted to download or to forward/distribute the text or part of it without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license (like Creative Commons).

The publication may also be distributed here under the terms of Article 25fa of the Dutch Copyright Act, indicated by the "Taverne" license. More information can be found on the University of Groningen website: <https://www.rug.nl/library/open-access/self-archiving-pure/taverne-amendment>.

### Take-down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Downloaded from the University of Groningen/UMCG research database (Pure): <http://www.rug.nl/research/portal>. For technical reasons the number of authors shown on this cover page is limited to 10 maximum.



## Chapter 3.1

# **What is the right blood hematocrit preparation procedure for standards and quality control samples for dried blood spot analysis?**

R.A. Koster  
J.W.C. Alffenaar  
R. Botma  
B. Greijdanus  
D.J. Touw  
D.R.A. Uges  
J.G.W. Kosterink

## **ABSTRACT**

Since hematocrit effects on dried blood spot analysis is widely acknowledged, a correct preparation of the target hematocrit is vital. We compared two procedures for preparing specific hematocrit values using a hematology analyzer. In the first procedure, whole blood was centrifuged, red blood cells (RBC) and plasma were separated and the proper volumes of RBC and plasma were mixed to obtain the target hematocrit. This procedure resulted in a measured hematocrit which was 11% lower than the target value. In the second procedure, blood was centrifuged and a calculated volume of plasma was removed or added in order to adjust the hematocrit. No difference was observed between the measured hematocrit and the target value. The second procedure was therefore considered the preferred procedure. It is recommended that as part of quality assurance the target hematocrit should be measured after preparation using a hematology analyzer.

## INTRODUCTION

Dried Blood Spot (DBS) sampling has many advantages and among others allows the patient to sample blood at home and send the DBS sample to the laboratory by mail. This sampling is simple; saves patients travel expenses and time, and is therefore considered to be patient friendly and cost effective. Other advantages of DBS sampling are lower risk of bio-hazard and a smaller amount of blood required for sampling, which is also an advantage for pre-clinical studies using rodents [1-4].

Several factors need to be taken into account in case therapeutic drug monitoring (TDM) using whole blood or plasma sampling is replaced by DBS sampling. The bias caused by the hematocrit (HT) variability of the sample can be considered a critical parameter related to DBS analysis [5]. The HT represents the relative volume of the red blood cells (RBC) in the blood and has a direct effect on the viscosity of the blood. The permeability of the DBS card is influenced by the HT of the blood. Blood with a high HT has a low penetration of the DBS card and thus forms a smaller spot. A fixed diameter punch would then contain a higher blood volume causing a higher bias. Earlier publications have investigated the effect on the measured concentration caused by the viscosity of blood over a range of HT values during validation. It is suggested to correct for this effect with the use of a linear relation between HT and measured concentration [6-8]. When a fixed volume of venous blood is drawn from the patient and applied on the DBS card, followed by full spot analysis, the DBS can be considered more as a sample storage method than as a sampling method. In this case the hematocrit effect caused by the viscosity of the blood would be of no influence on the analytical results. However, this would hamper self sampling by patients, because the applied volume cannot be checked for accuracy.

There are two procedures to prepare a target HT value of anticoagulated whole blood used for standard and quality control samples for DBS analysis. In the first procedure, anticoagulated blood is centrifuged, RBC and plasma are separated and the proper volumes of packed RBC and plasma are pipetted and pooled to obtain the target HT [9-15].

In the second procedure, a volume of anticoagulated blood with a known measured HT (by the microHT method or hematology analyzer) is centrifuged and the appropriate amount of plasma is removed or added to obtain the target HT [16-20]. The disadvantage of the second procedure is that the HT of each batch of blood needs to be measured by a hematology analyzer, but there's no tedious pipetting of the packed RBC. The first procedure seems to

control both the volume of RBC and the volume of plasma. However, the centrifuged red blood cell fraction still contains plasma between the cells. The volume of plasma which is still present in the red blood cell fraction may depend on the centrifugal procedure and may result in a lower HT value than the target HT value. Various centrifugal procedures may produce different HT values. In current literature, the HT preparation procedure is often not accurately described and may have had a negative influence on the target HT [7, 8, 15, 21-27]. Articles that describe the preparation of a target HT by pipetting RBC and plasma use varying centrifugal procedures like 2 minutes at 1000g [9], 10 minutes at 1543g [10, 12], 4 minutes at 7000g [28] and 10 minutes at 3506g [14]. It seems uncommon to refer to the DIN 58933-1 (German National Standard) reference method, the CLSI H7-A3 (Clinical Laboratory Standards Institute) standard or another reference method in the preparation of HT values for DBS analysis [29, 30]. The DIN 58933-1 document mentions that the product of the minimum relative centrifugal force ( $RCF \geq 5000g$ ) acting on the erythrocytes and the centrifuging time in minutes is at least 100,000, implying a centrifugal time for example of 20 minutes at 5000g ( $20 \times 5000 = 100,000$ ). The product of the centrifugal time and the RCF of 100,000 was not met by the earlier described centrifugal procedures for the preparation of the target HT by pipetting RBC and plasma [9-14], which ranged from 2000 (2 min x 1000g) [9] to 35060 (10 min x 3506g) [14]. The CLSI H7-A3 standard describes the use of a microhematocrit centrifuge spinning at 10,000g to 15,000g. The centrifuge method should be calibrated by testing the minimum packing time which is required to obtain a stable minimum packed cell volume for two consecutive time series [30]. For this study, the optimal centrifuge method was derived from both the DIN 58933-1 reference method and the CLSI H7-A3 standard. Since HT effects are common in DBS analysis, the preparation procedure of the target HT was considered a critical step in DBS analysis.

The varying HT preparation procedures described in literature may cause a variation in the prepared HT values and may influence the analytical results in DBS analysis. This effect may be underestimated at this time and was thus subject of this investigation.

The aim of the study was to establish the best procedure for the preparation of target HT values based on two common procedures. The first procedure involved pipetting RBC and plasma while the second procedure involved removing or adding plasma. These procedures were compared with some different HT preparation procedures described in literature and were all evaluated with the HT assay results of a hematology analyzer.

## MATERIALS AND METHODS

### Chemicals and materials

Citrate whole blood was purchased from Sanquin (Amsterdam, The Netherlands) and was used within two weeks after blood donation and was always checked on hemolysis before use. The whole blood was stored at 4°C. A Hettich centrifuge model 460R (Tuttlingen, Germany) was used to centrifuge the whole blood for HT preparation.

A XN-9000 hematology analyzer from Sysmex (Hyogo, Japan), using flow cytometry, was applied for all HT analyses and was an integral part of the routine patient HT analyses in the authors' hospital.

### The preparation of the target HT

In order to test the HT preparation procedure, HT calibration curves from three different lots were each prepared in triplicate by the two described different preparation procedures in order to achieve target HT values of 0.10, 0.20, 0.30, 0.40, 0.50, and 0.60 L/L. The whole blood was first allowed to reach ambient temperature under very gentle tumble mixing before any centrifugation tests were performed.

The first preparation procedure was performed by centrifuging citrate whole blood and separating the plasma supernatant and the RBC. Then, the RBC and plasma were mixed together in order to obtain the target HT. The RBC were pipetted using an Eppendorf air displacement pipette with two times pre-wetting of the pipette tip. Pre-wetting is a standard procedure for viscous samples recommended by Eppendorf AG (Hamburg, Germany), the producer of the Eppendorf pipette.

The second preparation consisted of centrifuging citrate whole blood with a known HT (measured by the Sysmex XN-9000 analyzer), followed by removing or adding the necessary volumes of plasma to achieve the target HT values. The two optimized HT preparation procedures are described in table 1.

Centrifugation procedures found in DBS publications, which make use of the pipetting of RBC and plasma, were used to prepare target HT values of 0.50 L/L in triplicate, as described above. The following centrifugation procedures were tested: 2 minutes at 1000g [9], 10 minutes at 1543g [10, 12], 4 minutes at 7000g [28] and 10 minutes at 3506g [14].

**Table 1 The two optimized hematocrit (HT) preparation procedures**

Step	First procedure	Second procedure
1	Centrifuge for 30 minutes at 4637g	Measure the HT with a hematology analyzer
2	Transfer plasma supernatant to fresh tube	Centrifuge for 5 minutes at 2000g
3	Discard approximately 300 µL of the top layer to minimize the amount of plasma in the RBC section	Remove or add the necessary volume of plasma
4	Gently homogenize the RBC	Gently homogenize the blood tube
5	Pipette the necessary volumes of RBC and plasma together	Measure the HT with a hematology analyzer
6	Measure the HT with a hematology analyzer	


The pipetting of the red blood cells (RBC) was performed by reverse pipetting with two times pre-wetting, using an Eppendorf air displacement pipette.

The prepared samples were analyzed for their HT using the Sysmex XN-9000. The prepared target HT and the measured HT of the HT calibration curves were compared using Passing and Bablok regression analysis with Analyse-it® (version 2.30) software (Leeds, United Kingdom).

## RESULTS AND DISCUSSION

The pipetting of the RBC appeared to be a critical step in the DBS procedure because of the high viscosity of the RBC. The use of a Gilson positive displacement pipette, which is very suitable for pipetting viscous samples, caused hemolysis by the friction of the plunger. An air displacement pipette (Eppendorf Hamburg, Germany) with careful reversed pipetting and two times pre-wetting proved to be the best approach. Centrifuging at 4637g for 30 minutes resulted in no further decrease in RBC levels compared with the preceding centrifugation time, which complied with the CLSI H7-A3 standard. The centrifugal procedure also complied with the DIN 58933-1 guideline, where the RCF multiplied with the time in minutes, was in this case 138,000 and was higher than the recommended value of >100,000 (where the RCF should be >5000).

For the procedure of pipetting the RBC and plasma, the Passing and Bablok results of the three prepared HT curves in triplicate showed a difference of -11% between measured HT relative to target HT according to the following equation:  $y = 0.89x - 0.01$ , with 95% confidence intervals (CI's) of 0.86 to 0.95 for the slope and -0.02 to 0.01 for the intercept (figure 1).




**Figure 1** Results of the Passing and Bablok and residual plots where the prepared target hematocrit was compared with the measured hematocrit using a Sysmex XN-9000 analyzer.

The target hematocrit was prepared by pipetting red blood cells and plasma in triplicate for each of the 3 batches of blood at each hematocrit level.  $N = 9$  at each of the 6 hematocrit levels. The following equation was found:  $y = 0.89x - 0.01$ , with 95% confidence intervals (CI's) of 0.86 to 0.95 for the slope and -0.02 to 0.01 for the intercept.

The coefficients of variation (CV's) for the HT measurements ranged between 3.95% and 7.04% per HT value ( $n=9$ ). The procedure of adding or omitting plasma from blood with a measured HT showed a deviation in HT of 0% according to the following equation:  $y = 1.00x - 0.00$ , with 95% CI's of 0.95 to 1.03 for the slope and -0.01 to 0.01 for the intercept (figure 2).





**Figure 2** Results of the Passing and Bablok and residual plots where the prepared target hematocrit was compared with the measured hematocrit using a Sysmex XN-9000 analyzer.

The target hematocrit was prepared by removing or adding plasma in triplicate for each of the 3 batches of blood at each hematocrit level.  $N = 9$  at each of the 6 hematocrit levels. The following equation was found:  $y = 1.00x - 0.00$ , with 95% CI's of 0.95 to 1.03 for the slope and -0.01 to 0.01 for the intercept.

The CV's for the HT measurements ranged between 2.13% and 3.65% per HT value ( $n=9$ ). The procedure of adding or omitting plasma yielded lower CV's compared to the pipetting of RBC and plasma, indicating that this is a more robust procedure.

The varying centrifugal times and forces combined with pipetting of the RBC and plasma as mentioned in previous publications provided similar deviations of the HT, as can be seen in table 2.

The procedure of centrifuging for 2 minutes at 1000g showed by far the highest deviation, and can be considered as inadequate. The differences between the other centrifugal procedures were of much less influence on the HT preparation procedure. However, all the centrifugal procedures that used the pipetting of RBC and plasma yielded biases between -7.9% and -16.1%. It can be concluded that the procedure of pipetting RBC and plasma will unavoidably cause an error in the prepared HT. And this procedure should only be used after correction by HT measurements for every HT preparation using a hematology analyzer. Maybe this procedure could be validated over time with various batches of blood, where a constant bias and a low batch to batch variability may provide a correction factor.

As a consequence of these discrepancies, the HT effects mentioned in the literature should be interpreted with caution, since real HT values may deviate from the mentioned HT values. A HT preparation procedure that causes deviating HT values is especially advised against when a HT effect has been observed and standards and quality control samples are prepared with a HT value that was meant to represent the patient population. This prepared HT value may already cause analytical biases due to the different HT values of the standards and the patient population.

**Table 2 The varying centrifugal times and forces combined with pipetting of the RBC and plasma derived from previous publications**

Centrifugal parameters	Bias (%)	CV (%)
2 minutes at 1000g	-16.1	0.14
10 minutes at 1543g	-12.1	0.57
10 minutes at 3506g	-9.7	0.78
5 minutes at 10,000g	-9.2	1.14
30 minutes at 4637g	-9.1	1.25
4 minutes at 7000g	-7.9	0.13

The prepared target hematocrit values were compared to the hematology analyzer resulting in the stated biases (n=3).

## CONCLUSIONS

The preparation of the target HT with the pipetting of RBC and plasma (first procedure) caused a significant bias in the HT and can only be used when the HT is corrected with the measured HT using a hematology analyzer. The procedure of adding or removing plasma from blood with a known HT (second procedure) caused no error in the HT and was considered the preferred HT preparation procedure. It is recommended that as part of quality assurance the target HT should be measured after preparation using a hematology analyzer.

## FUTURE PERSPECTIVE

The use of dried blood spot (DBS) analysis for therapeutic drug monitoring will play an increasing role in the future. The investigation of critical parameters such as the preparation of the target hematocrit value for calibration curves and quality control samples will lead to improved procedures for DBS validation and analysis. This will ensure that future DBS analysis results will be generated with improved quality standards.

## REFERENCES

1. Li W, Tse FL. Dried blood spot sampling in combination with LC-MS/MS for quantitative analysis of small molecules. *Biomed. Chromatogr.* 24(1), 49-65 (2010).
2. Vu DH, Alffenaar JW, Edelbroek PM, Brouwers JR, Uges DR. Dried blood spots: a new tool for tuberculosis treatment optimization. *Curr. Pharm. Des.* 17(27), 2931-2939 (2011).
3. Edelbroek PM, van der Heijden J, Stolk LM. Dried blood spot methods in therapeutic drug monitoring: methods, assays, and pitfalls. *Ther. Drug Monit.* 31(3), 327-336 (2009).
4. Chapman K, Chivers S, Gliddon D et al. Overcoming the barriers to the uptake of nonclinical microsampling in regulatory safety studies. *Drug Discov. Today.* 19(5), 528-532 (2014).
5. Wilhelm AJ, den Burger JC, Swart EL. Therapeutic Drug Monitoring by Dried Blood Spot: Progress to Date and Future Directions. *Clin. Pharmacokinet.* (2014).
6. Koster RA, Alffenaar JW, Greijdanus B, Uges DR. Fast LC-MS/MS analysis of tacrolimus, sirolimus, everolimus and cyclosporin A in dried blood spots and the influence of the hematocrit and immunosuppressant concentration on recovery. *Talanta.* 115, 47-54 (2013).
7. Vu DH, Koster RA, Alffenaar JW, Brouwers JR, Uges DR. Determination of moxifloxacin in dried blood spots using LC-MS/MS and the impact of the hematocrit and blood volume. *J. Chromatogr. B. Analyt. Technol. Biomed. Life. Sci.* 879(15-16), 1063-1070 (2011).
8. den Burger JC, Wilhelm AJ, Chahbouni A, Vos RM, Sinjewel A, Swart EL. Analysis of cyclosporin A, tacrolimus, sirolimus, and everolimus in dried blood spot samples using liquid chromatography tandem mass spectrometry. *Anal. Bioanal. Chem.* (2012).

9. Moller I, Thomas A, Geyer H, Schanzer W, Thevis M. Development and validation of a mass spectrometric detection method of peginesatide in dried blood spots for sports drug testing. *Anal. Bioanal Chem.* 403(9), 2715-2724 (2012).
10. Mess JN, Taillon MP, Cote C, Garofolo F. Dried blood spot on-card derivatization: an alternative form of sample handling to overcome the instability of thiorphan in biological matrix. *Biomed. Chromatogr.* 26(12), 1617-1624 (2012).
11. Patel AJ, Wesley R, Leitman SF, Bryant BJ. Capillary versus venous haemoglobin determination in the assessment of healthy blood donors. *Vox Sang.* 104(4), 317-323 (2013).
12. Youhnovski N, Bergeron A, Furtado M, Garofolo F. Pre-cut dried blood spot (PCDBS): an alternative to dried blood spot (DBS) technique to overcome hematocrit impact. *Rapid Commun. Mass Spectrom.* 25(19), 2951-2958 (2011).
13. Kvaskoff D, Ko P, Simila HA, Eyles DW. Distribution of 25-hydroxyvitamin D3 in dried blood spots and implications for its quantitation by tandem mass spectrometry. *J. Chromatogr. B. Analyt Technol. Biomed. Life. Sci.* 901, 47-52 (2012).
14. Vu DH, Koster RA, Bolhuis MS et al. Simultaneous determination of rifampicin, clarithromycin and their metabolites in dried blood spots using LC-MS/MS. *Talanta.* 121, 9-17 (2014).
15. Rhoden L, Antunes MV, Hidalgo P, Silva CA, Linden R. Simple procedure for determination of valproic acid in dried blood spots by gas chromatography-mass spectrometry. *J. Pharm. Biomed. Anal.* 96, 207-212 (2014).
16. Cox HD, Rampton J, Eichner D. Quantification of insulin-like growth factor-1 in dried blood spots for detection of growth hormone abuse in sport. *Anal. Bioanal Chem.* 405(6), 1949-1958 (2013).
17. O'Mara M, Hudson-Curtis B, Olson K, Yueh Y, Dunn J, Spooner N. The effect of hematocrit and punch location on assay bias during quantitative bioanalysis of dried blood spot samples. *Bioanalysis.* 3(20), 2335-2347 (2011).
18. Zheng JH, Guida LA, Rower C et al. Quantitation of tenofovir and emtricitabine in dried blood spots (DBS) with LC-MS/MS. *J. Pharm. Biomed. Anal.* 88, 144-151 (2014).
19. Shah NM, Hawwa AF, Millership JS, Collier PS, McElnay JC. A simple bioanalytical method for the quantification of antiepileptic drugs in dried blood spots. *J. Chromatogr. B. Analyt Technol. Biomed. Life. Sci.* 923-924, 65-73 (2013).
20. Jimmerson LC, Zheng JH, Bushman LR, MacBrayne CE, Anderson PL, Kiser JJ. Development and validation of a dried blood spot assay for the quantification of ribavirin using liquid chromatography coupled to mass spectrometry. *J. Chromatogr. B. Analyt Technol. Biomed. Life. Sci.* 944, 18-24 (2014).
21. Rao RN, Raju SS, Vali RM, Sankar GG. Liquid chromatography-mass spectrometric determination of losartan and its active metabolite on dried blood spots. *J. Chromatogr. B. Analyt Technol. Biomed. Life. Sci.* 902, 47-54 (2012).
22. Rao RN, Maurya PK, Ramesh M, Srinivas R, Agwane SB. Development of a validated high-throughput LC-ESI-MS method for determination of sirolimus on dried blood spots. *Biomed. Chromatogr.* 24(12), 1356-1364 (2010).
23. Wilhelm AJ, den Burger JC, Vos RM, Chahbouni A, Sinjewel A. Analysis of cyclosporin A in dried blood spots using liquid chromatography tandem mass spectrometry. *J. Chromatogr. B. Analyt Technol. Biomed. Life. Sci.* 877(14-15), 1595-1598 (2009).

24. Hinchliffe E, Adaway JE, Keevil BG. Simultaneous measurement of cyclosporin A and tacrolimus from dried blood spots by ultra high performance liquid chromatography tandem mass spectrometry. *J. Chromatogr. B. Analyt Technol. Biomed. Life. Sci.* (2011).
25. van der Heijden J, de Beer Y, Hoogtanders K et al. Therapeutic drug monitoring of everolimus using the dried blood spot method in combination with liquid chromatography-mass spectrometry. *J. Pharm. Biomed. Anal.* 50(4), 664-670 (2009).
26. Koop DR, Bleyle LA, Munar M, Cherala G, Al-Uzri A. Analysis of tacrolimus and creatinine from a single dried blood spot using liquid chromatography tandem mass spectrometry. *J. Chromatogr. B. Analyt Technol. Biomed. Life. Sci.* 926, 54-61 (2013).
27. Della Bona ML, Malvagia S, Villanelli F et al. A rapid liquid chromatography tandem mass spectrometry-based method for measuring propranolol on dried blood spots. *J. Pharm. Biomed. Anal.* 78-79, 34-38 (2013).
28. Patel P, Tanna S, Mulla H, Kairamkonda V, Pandya H, Lawson G. Dexamethasone quantification in dried blood spot samples using LC-MS: The potential for application to neonatal pharmacokinetic studies. *J. Chromatogr. B. Analyt Technol. Biomed. Life. Sci.* 878(31), 3277-3282 (2010).
29. DIN Deutsches Institut für Normung. Bestimmung des Volumenanteils der Erythrocyten im Blut. DIN, Berlin (1995).
30. NCCLS. Procedure for Determining Packed Cell Volume by the Microhematocrit Method; Approved Standard. NCCLS document H7-A3 [ISBN 1-56238-413-9] (2000).



